

校准证书

CALIBRATION CERTIFICATE

证书编号: [2020DW02330333]

Certificate No.

委托方名称 Customer	杭州市地铁集团有限责任公司运营分公司
委托方地址 Address	杭州市江干区德胜东路5277号
样品名称 Name Of Sample	直流高压发生器
制造厂商 Manufacturer	武汉国电华美电气设备有限公司
型号规格 Model/Type	HMZGF-120kV/2mA
器具编号 No Of Sample	2007102

证书专用章

Stamp

湖北省计量测试技术研究院
证书骑缝章



校准日期
Calibration date

2020 年 08 月 28 日
Y M D

证书批准人
Approved by

核验员
Checked by

校准员
Calibrated by

耿睿

李亮波

田天

本次校准所使用的测量装置均溯源至保存在中国计量科学研究院的国家计量基准。中国计量科学研究院于1999年代表中国签署了国际间“国家计量基准及国家计量研究院出具的校准和测量证书相互承认协议”。

The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.



● 本院是政府计量行政管理部门依法设立的法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

● 本院质量管理体系符合ISO/IEC17025标准的要求。

The quality management system for laboratory complies with ISO/IEC 17025 standards.

● 本次校准的技术依据 (名称、代号)

Reference documents for the Calibration (Name、Code)

参照: HJ/QC 24002-2015 直流高压发生器校准方法

● 本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

设备名称

Name of Equipment

200kV直流标准电阻分压器

数字多用表

型号/编号

Model/Serial No.

GF-200/251185

34401A/MY45016530

证书号/有效期

Certificate No./Due Date

(计)字

第201940043号/2021-01-13

2020DW02310496/2021-06-17

● 校准环境条件

Environmental condition on the Calibration

温度: 26.5°C

Temperature

气压: ———

Pressure

相对湿度: 46 %

R.H.

地点: 水果湖基地高压室

Place

其它: ———

Others

原始记录编号: 2020DW02330333

Record No.

本校准结论, 仅对受校样品的本次校准有效。

It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated.

未经本院许可, 不得部分复制本证书。

校准数据/结果

Data/Results of Calibration

1、直流高压测量:		
显示值 (kV)	实测标准值 (kV)	测量不确定度 $U_{rel}, k=2$
20.0	20.31	0.3%
40.0	40.60	0.2%
60.0	60.69	0.2%
80.0	80.65	0.2%
100.0	100.38	0.2%
120.0	120.14	0.2%
2、直流电流测量 (显示值为微安表读数):		
显示值 (μA)	实测标准值 (μA)	测量不确定度 $U_{rel}, k=2$
648	646.1	0.5%
1101	1101.6	0.5%
1947	1929.6	0.5%
注: 建议再校期为一年。		

以下空白